



校准证书

CALIBRATION CERTIFICATE

校准证书编号: YFJC/01-第 2018-003813 号

Calibrated certificate series No.

委 托 者 常州仁成金属制品有限公司
 Customer

委 托 者 地 址 /
 Address of customer

器 具 名 称 外径千分尺
 Name of instrument

制 造 厂 /
 Manufacturer

型 号 / 规 格 (25-50) mm
 Model/Specification

器 具 编 号 CZRC-002
 NO. of instrument



证书批准人 [Signature]
 Approved by

核 验 员 [Signature]
 Checked by

校 准 员 [Signature]
 Calibrated by

校准日期 2018 年 10 月 18 日
 Date for calibrated Year Month Day

本证书提供的测量结果仅对本次被校对象有效。 The result of the certificate is effect only on this calibrated instrument (s).
 未经许可，不得部分复制本证书。 This certificate can not be partly if not allowed.

地址: 上海市闵行区北翟路 1551 号 Address: Shanghai Minhang district Beidai road No 1551
 邮政编码: 201100 Post Code: 201100
 电话: 62388221 Tel: 62388221
 传真: 52191183 Fax: 52191183





上海英菲计量设备检测有限公司

SHANGHAIYINGFEI MEASUREMENTEQUIPMENT EXAMINATIONCO.,LTD.

校准证书编号: YFJC/01-第 2018-003813 号

Calibrated certificate series No

上海市计量校准实验室认可证书号: SJ052

The number of the certificate accredited by Shanghai Laboratory of Measurement and Calibration is No.

本次校准所依据的技术规范(代号、名称):

Reference documents for the calibration(code, name)

JJG21-2008 (千分尺检定规程)

YFJC/BQDD-01-02 (测微类量具示值误差测量结果不确定度评定)

本次校准所使用的主要标准器具:

Main measurement standards used in this calibration

名称/型号 Name/Model	编号 Number	证书编号/有效期限 Certificate No./Due date	测量范围/准确度 Measuring range/accuracy
量块/20 块组	10692	2018D10-20-000529/ 2019.05.28	(5.12~100)mm 4 等

出证机构:上海市计量测试技术研究院

Issuing body:Shanghai metrological testing technology research institute

以上计量标准器具的量值溯源至国家基准。

Quantity values of above measurement standards used in this calibration are traced to those of the national primary standards in the P.R.China.

校准地点及环境条件:

Location and environmental condition for the calibration

地点: 现场

Location

温度: 22℃;

Ambient temperature

湿度: 58%RH;

Relative humidity

其它: /

Others

本次校准结果的扩展不确定度:

Expanded uncertainty

$U=1.6 \mu m$ (k=2)

被校器具限制使用范围及条件:

The limitory scope and condition

校准证书续页专用

Continued page of calibration certificate



第 2 页 共 3 页

Page of total pages



上海英菲计量设备检测有限公司

SHANGHAIYINGFEI MEASUREMENTEQUIPMENT EXAMINATIONCO.,LTD.

校准证书编号:YFJC/01-第2018-003813号

Calibrated certificate series No.

校准结果/说明:

Results of calibration and additional explanation

受检点 (mm)	测量值 (μm)	量块偏差 (μm)	示值误差 (μm)
A+5.12	-1	-0.10	-1
A+10.24	-1	-0.18	-1
A+15.36	-1	-0.24	-1
A+21.50	0	0.08	0
A+25.00	1	0.26	1

本次校准内容结束

经校准符合技术要求

建议: 下次计量校准日期为2019年10月17日

校准证书续页专用

Continued page of calibration certificate



第 3 页 共 3 页

Page of total pages